

AIR SAMPLING ANALYSIS

THIS IS TO CERTIFY AN AIR SAMPLE SUPPLIED BY

Hyperbaric Technologies, Inc. - Portable Hyperbaric Chamber

was obtained by **RPT Safety and Health Services, LLC**, Glenville, New York, using an Evacuated Canister sample for an aromatic hydrocarbon profile and volatile organics profile including a library search for tentatively identified compounds using analytical method - mod. OSHA PV2120/TO15:

Results of all samples were less than the level of quantitation.

(Substance and sample results are listed on the reverse side of this certificate)

No other representations or warranties are expressed or implied other than the analysis results listed below.



**Richard M. Crossman, OHST
Analyst
December 16, 2005**

**RPT Safety and Health Services, LLC.
16 Harmon Road, Scotia, New York 12302
518.384.1365**

Evacuated CanisterAromatic Hydrocarbon Profile and Volatile Organics Profile.
Library Search for Tentatively Identified Compounds

Substance	Detection Limit - ppbv	Sample Results - ppmv
m & p- xylene	100	<100
Styrene	50	<50
0-Xylene	50	<50
1,1,2,2-Tetrachloroethane	50	<50
4-Ethyltoluene	50	<50
1,3,5-Trimethylbenzene	50	<50
1,2,4-Trimethylbenzene	50	<50
1,3-Dichlorobenzene	50	<50
Benzyl Chloride	50	<50
1,4-Dichlorobenzene	50	<50
1,2-Dichlorobenzene	50	<50
1,2,4-Trichlorobenzene	50	<50
Hexachloro-1,3-butadiene	50	<50
Chloroform	50	<50
Tetrahydrofuran	50	<50
1,2-Dichloroethane	50	<50
1,1,1-Trichloroethane	50	<50
Cyclohexane	50	<50
Carbon Tetrachloride	50	<50
Benzene	50	<50
1,1-Dioxane	200	<200
2,2,4-Trimethylpentane	50	<50
Heptane	50	<50
1,2,-Dichloropropane	50	<50
Trichloroethane	50	<50
Bromodichloromethane	50	<50
Cis-1,3-Dichloropropene	50	<50
Trans-1,3-Dichloropropene	50	<50
Dichloropropene		
1,1,2-Trichloroethane	50	<50
Dibromochloromethane	50	<50
Methyl Isobutyl Ketone	200	<200
Methyl Butyl Ketone	200	<200
1,2,-Dibromoethane	50	<50
Tetrachloroethane	50	<50
Chlorobenzene	50	<50
Ethylbenzene	50	<50
Bromoform	50	<50
Propylene	50	<50
Freon-12	50	<50
Chloromethane	50	<50
Freon-114	50	<50
Vinyl Chloride	50	<50
1,3,-Butadiene	50	<50
Chloroethane	50	<50
Vinyl Bromide	50	<50
Bromomethane	50	<50
Freon-11	50	<50
Isopropyl Alcohol	50	<50
Acetone	50	<50
1,1,-Dichloroethane	50	<50
Methylene Chloride	50	<50
Freon-113	50	<50
Allyl Chloride	50	<50
Carbon Disulfide	50	<50
Trans-1,1-Dichloroethane	50	<50
Methyl Tert-Butyl ether	50	<50
1,1,Dichloroethene	50	<50
Vinyl Acetate	50	<50
Methyl Ethyl Ketone	50	<50
Cis-1,2-Dichloroethylene	50	<50
Hexane	50	<50
Ethyl Acetate	50	<50